






QUALICHEK™ Clenbuterol ELISA

REF : KBFP1012

Ver 1.0

RUO

Quantitative testing of Clenbuterol in samples, such as urine, tissues and feed, etc.

RUO	For Research Use Only	REF	Catalog Number
	Store At	LOT	Batch Code
	Manufactured By		Biological Risk
	Expiry Date		Consult Operating Instructions

For Research Purposes Only Purchase does not include or carry the right to resell or transfer this product either as a stand-alone product or as a component of another product. Any use of this product other than the permitted use without the express written authorization of KRISHGEN BioSystems is strictly prohibited.



KRISHGEN BioSystems | For US / Europe: toll free +1(888)-970-0827 tel: +1(562)-568-5005
 For Asia / India: tel: +91(22)-49198700
 Email: sales@krishgen.com

Introduction:

Clenbuterol is a sympathomimetic amine used by sufferers of breathing disorders as a decongestant and bronchodilator. People with chronic breathing disorders such as asthma use this as a bronchodilator to make breathing easier. It is most commonly available as the hydrochloride salt, clenbuterol hydrochloride.

Intended Use:

This kit is for quantitative testing for Clenbuterol in sample, such as urine, tissues and feed etc.

Principle:

This kit uses Indirect-Competitive-ELISA as the method. It can detect CLE in samples, such as urine, tissues and feed etc. This kit is composed of ELISA Microtiter plate, HRP conjugate, antibody, standard and other supplementary reagents. The microplate provided in this kit has been pre-coated with coupled antigen. During the reaction, CLE in the samples or standard competes with coupled antigen on the solid phase supporter for sites of CLE antibody. Then Horseradish Peroxidase (HRP) conjugate is added to each micro plate well, and TMB substrate is for color development. There is a negative correlation between the OD value of samples and the concentration of CLE. The concentration of CLE in the samples can be calculated by comparing the OD of the samples to the standard curve.

Materials Provided:

1. Coated Microtiter plate - 1x 96 wells
2. Standards Liquid -1mL each (0 ppb, 0.1 ppb, 0.3 ppb, 0.9 ppb, 2.7 ppb, 8.1 ppb)
3. High Concentrated Standard (100 ppb) - 1mL
4. HRP Conjugate - 5.5 mL
5. Antibody Working Solution - 9mL
6. Substrate A - 6 mL
7. Substrate B - 6 mL
8. Stop Solution - 6 mL
9. Wash Buffer (20X) - 40 mL
10. Plate Sealer - 3 pieces

Material required but not provided:

Instrument: Microtiter plate reader, Printer, Homogenizer, Nitrogen Evaporators/Water bath, Oscillators, centrifuge, Graduated pipette, Balance (sensitivity 0.01 g)

High-precision transferpettor: single channel (20-200 μ L, 100-1000 μ L), Multichannel (300 μ L) micropipettes.

Reagents: NaOH, Acetic ether, Concentrated HCl, Acetonitrile, Methanol, N-hexane, Anhydrous sodium sulfate.

Storage Information:

1. Store main kit components at 2-8°C Avoid freeze / thaw cycles.
2. Before using, bring all components to room temperature (18-25°C). Upon assay completion return all components to appropriate storage conditions

Sample pretreatment:

Bring all reagents and samples to room temperature before use.

Open the microplate reader in advance, preheat the instrument, and set the testing parameters.

1. Sample pretreatment Notice:

Experimental apparatus should be clean, and the pipette should be disposable to avoid cross-contamination during the experiment.

2. Solution preparation:

Solution 1: 0.1 M HCl Solution: Dilute 0.86 mL Concentrated HCl with deionized water to 100 mL.

Solution 2: 0.1 M NaOH Solution: Dissolve 0.4 g NaOH with 100 mL deionized water

Solution 3: Acetonitrile-0.1 M HCl Solution: Volume (Acetonitrile): Volume (0.1 M HCl solution) =84:16.

Solution 4: Reconstitution Buffer: Dilute the 10×Reconstitution Buffer with deionized water. (10×Reconstitution Buffer (V): Deionized water (V)=1:9). The Reconstitution buffer can be store at 4°C for a month.

Solution 5: Wash Buffer.: Dilute 20×Concentrated Wash Buffer with deionized water (20×Concentrated Wash Buffer (V): Deionized water (V) = 1:19).

3. Sample pretreatment procedure:**Pretreatment of urine sample:**

Take 20 µL of clear urine sample for analysis directly (if the urine sample is muddy, it should be filtered or centrifuged at 4000 r/min for 5 min until the urine sample become clear). Samples temporarily not used should be frozen.

Note: Sample dilution factor: 1, minimum detection dose: 0.1 ppb.

Pretreatment of tissue (pretreatment method 1):

Weigh 2±0.05 g of crushed homogenate tissue sample, add 6 mL of Reconstitution buffer. Oscillate fully for 2 min, centrifuge at a speed of over 4000 r/min for 10 min (incubate the sample at 85°C for 10 min before centrifugation if there is a high-content of fat in tissue sample). Take 20 µL of the supernatant for analysis.

Note: Sample dilution factor: 4, minimum detection dose: 0.4 ppb.

Pretreatment of tissue (pretreatment method 2):

1. Weigh 2±0.05 g of crushed homogenate tissue sample, add 6 mL of acetonitrile-0.1M HCl solution. Oscillate for 2 min, centrifuge at a speed of over 4000 r/min for 10 min at room temperature.
2. Take 3 mL of the supernatant. Add 2 mL of 0.1 M NaOH and 6 mL of acetic ether. Oscillate fully for 2 min, centrifuge at a speed of over 4000 r/min for 10 min at room temperature. Take all the supernatant and dry at 50-60°C with nitrogen or Water bath.
3. Add 1 mL of Reconstitution buffer and oscillate for 30 sec. Take 20 µL for analysis.

Note: Sample dilution factor: 1, minimum detection dose: 0.1 ppb.

Pretreatment of feed sample:

1. Weigh 1±0.05 g of homogenate feed sample, add 10 mL of methanol and 5 g of Na₂SO₄. Oscillate for 2 min, centrifuge at a speed of over 4000 r/min for 10 min at room temperature.
2. Take 1 mL of the supernatant and dry with nitrogen or Water bath at 50-60°C. Add 1 mL of Reconstitution buffer to dissolve the remaining dry material. Then add 1 mL of n-hexane and mix for 30 sec. Centrifuge for 5 min at a speed of over 4000 r/min at room temperature.
3. Take 20 µL of the lower layer liquid for analysis.

Note: Sample dilution factor: 10, minimum detection dose: 1 ppb.

Assay procedure:

Centrifuge the sample again after thawing before the assay. Bring all reagents to room temperature for 30 min before use. **All the reagents should be mixed thoroughly by gently swirling before pipetting. Avoid foaming.**

1. **Number:** number the sample and standard in order (multiple well), and keep a record of standard wells and sample wells.
2. **Add sample:** add 20 µL of Standard or Sample per well, then add 50µL of HRP conjugate to each well, then add 80 µL of antibody working solution, cover the plate with sealer, oscillate for 5 s gently to mix thoroughly, incubate for 30min at 25°C.
3. **Wash:** uncover the sealer carefully; remove the liquid in each well. Immediately add 300 µL of wash buffer to each well and wash. Repeat wash procedure for 5 times, 30 s intervals/time. Invert the plate and pat it against thick clean absorbent paper (If bubbles exist in the wells, clean tips can be used to prick them).
4. **Color Development:** add 50 µL of Substrate Reagent A to each well, and then add 50 µL of Substrate Reagent B. Gently oscillate for 5 s to mix thoroughly. Incubate shading light for 15 min at 25°C (The reaction time may be shortened or prolonged according to the depth of the color).
5. **Stop reaction:** add 50 µL of stop solution to each well, oscillate gently to mix thoroughly.
6. **OD Measurement:** determine the optical density (OD value) of each well at 450 nm with a microplate reader (the 450/630 nm double wavelength is recommended). This step should be finished in 10 min after stop reaction.

Judgment of result:**Absorbance (%) = $A/A_0 \times 100\%$**

A: Average absorbance of standard or sample

A₀: Average absorbance of 0 ppb Standard

Drawing and calculation of standard curve

Create a standard curve by plotting the absorbance percentage of each standard on the y-axis against the log concentration on the x-axis to draw a semi-logarithmic plot. Add average absorbance value of sample to standard curve to get corresponding concentration. If samples have been diluted, the concentration calculated from the standard curve must be multiplied by the dilution factor.

For this kit, it is more convenient to use professional analysis software for accurate and fast analysis on a large number of samples.

Notes:

1. The overall OD value will be lower when reagents have not been brought to room temperature before use or room temperature is below 25°C.
2. If the wells turn dry during the washing procedure, it will lead to bad linear standard curve and poor repeatability. Operate the next step immediately after wash.
3. Mix thoroughly and wash the plate completely. The consistency of wash procedure can strongly affect the reproducibility of this ELISA kit.
4. ELISA Microtiter plate should be covered by plate sealer. Avoid the reagents to strong light.
5. Do not use expired kit or reagents of different batches.
6. TMB should be abandoned if it turns color. When OD value of standard (concentration: 0) < 0.5 unit (A_{450nm} < 0.5), it indicates the reagent may be deteriorated.
7. Stop solution is caustic; avoid contact with skin and eyes.

Technical specification:**Sensitivity:** 0.1 ppb (ng/mL)**Reaction mode:** 25°C, 30 min~15 min**Detection limit:** Urine ---0.1 ppb, Tissue (treatment method 1) ---0.4 ppb,
Tissue (treatment method 2) ---0.1 ppb, Feed---1 ppb.**Reactivity:** Clenbuterol ---100%, Arubendol---<1%, Mabuterol---<1%,
Brombuterol---<1%, Albuterol---<1%, Ractopamine---<1%.**Sample recovery rate:** Urine ---95%±10%, Tissue, Feed---85%±15%.**LIMITED WARRANTY**

Krishgen Biosystems does not warrant against damages or defects arising in shipping or handling, or out of accident or improper or abnormal use of the product; against defects in products or components not manufactured by Krishgen Biosystems, or against damages resulting from such non-Krishgen Biosystems made products or components. Krishgen Biosystems passes on to customer the warranty it received (if any) from the maker thereof of such non-Krishgen made products or components. This warranty also does not apply to product to which changes or modifications have been made or attempted by persons other than pursuant to written authorization by Krishgen Biosystems.

THIS WARRANTY IS EXCLUSIVE. The sole and exclusive obligation of Krishgen Biosystems shall be to repair or replace the defective product in the manner and for the period provided above. Krishgen Biosystems shall not have any other obligation with respect to the products or any part thereof, whether based on contract, tort, strict liability or otherwise. Under no circumstances, whether based on this Limited Warranty or otherwise, shall Krishgen Biosystems be liable for incidental, special, or consequential damages.

This Limited Warranty states the entire obligation of Krishgen Biosystems with respect to the product. If any part of this Limited Warranty is determined to be void or illegal, the remainder shall remain in full force and effect.

Krishgen Biosystems. 2020

THANK YOU FOR USING KRISHGEN PRODUCT!