






KRIBIOLISA™ Ofatumumab (ARZERRA) ELISA

REF : KBI1516

Ver.1.0

RUO

Enzyme Immunoassay for the Quantitative Determination of
Ofatumumab in Human Serum and Plasma

RUO	For Research Use Only	REF	Catalog Number
	Store At	LOT	Batch Code
	Manufactured By		Biological Risk
	Expiry Date		Consult Operating Instructions

For Research Use Only. Purchase does not include or carry the right to resell or transfer this product either as a stand-alone product or as a component of another product. Any use of this product other than the permitted use without the express written authorization of KRISHGEN BioSystems is strictly prohibited.

REF KBI1516

 96 tests



KRISHGEN BioSystems |

For US / Europe: toll free +1(888)-970-0827 tel: +1(562)-568-5005
For Asia / India: tel: +91(22)-49198700
Email: sales1@krishgen.com

Introduction:

Ofatumumab is a fully human monoclonal antibody to CD20, which appears to provide rapid B-cell depletion. Under the brand name Kesimpta, it is approved for the treatment of multiple sclerosis in the United States as well as in the European Union and other regions. Under the brand name Arzerra, it is approved for the treatment of certain types of chronic lymphocytic leukemia (CLL).

Intended Use:

The KRIBIOLISA™ Ofatumumab (ARZERRA) ELISA kit is used for the quantitative estimation of Ofatumumab in human serum and plasma.

Principle:

The Ofatumumab ELISA is a competitive immunoassay for the determination of Ofatumumab. It is known that Ofatumumab binds strongly to Claudin-18 antigen. Hence using this principle the assay has been developed. A varying concentration of unlabeled standard or sample and constant concentration of Ofatumumab:HRP conjugate will bind in sequence to the antibodies coated on the microplate. Upon washing, unbound Ofatumumab:HRP Conjugate will be removed. Bound Ofatumumab:HRP complex will produce a soluble blue colored product after the addition of TMB Substrate. The enzyme reaction is stopped by dispensing of stop solution into the wells. The optical density (OD) of the solution at 450 nm is inversely proportional to the amount of bound Ofatumumab present in the standards or samples.

Materials Provided:

Part	Description	Qty
CLAUDIN-18 protein Coated Microtiter Plate	96 well polystyrene microplate (12 strips of 8 wells) coated with CLAUDIN-18 protein Coated Microtiter Plate	1 x 96 wells
Ofatumumab Standard	Ofatumumab in Buffered protein base with preservative thiomersol < 0.01% standard (lyophilized, concentration - 8,000 ng/ml)	2 vials
Ofatumumab:HRP Conjugate concentrated	Ofatumumab:HRP conjugate (concentrated 1mg/ml)	1 vial
Detection Diluent	Buffered protein base with protein stabilizer and preservatives 0.02% methylisothiazolone and 0.02% bromonitrodioxane.	12 ml
(1X) Sample Diluent	Buffered protein base with preservative thiomersol < 0.01%	2 x 50 ml
(1X) Standard Diluent	Buffered protein base with preservative thiomersol < 0.01% and 1:1000 dilution human serum	10 ml
(20X) Wash Buffer	20-fold concentrated solution of buffered surfactant with preservative thiomersol < 0.01%. May turn yellow over time.	25 ml
TMB Substrate	Stabilized TMB Chromogen	12 ml
Stop Solution	0.73M Phosphoric Acid	12 ml
Instruction Manual		1 no

Materials to be provided by the End-User:

1. Microplate Reader able to measure absorbance at 450 nm.
2. Adjustable pipettes to measure volumes ranging from 50 ul to 1000 ul.
3. Deionized (DI) water.
4. Wash bottle or automated microplate washer.
5. Graph paper or software for data analysis.
6. Tubes to prepare standard/sample dilutions.
7. Timer.
8. Absorbent paper.
9. Incubator

Handling/Storage:

1. All reagents should be stored at 2°C to 8°C for stability.
2. All the reagents and wash solutions should be used within 12 months from manufacturing date.
3. Before using, bring all components to room temperature (18-25°C). Upon assay completion ensure all components of the kit are returned to appropriate storage conditions.
4. The Substrate is light-sensitive and should be protected from direct sunlight or UV sources.

Health Hazard Warnings:

1. Reagents that contain preservatives may be harmful if ingested, inhaled or absorbed through the skin.
2. For Research Use Only.

**Sample Preparation and Storage:**

Blood is taken by venipuncture. Serum is separated after clotting by centrifugation. Plasma can be used, too. Lipaemic, hemolytic or contaminated samples should not be run. Repeated freezing and thawing should be avoided. If samples are to be used for several assays, initially aliquot samples and keep at -20°C.

For Cell Culture Supernatant – If necessary, centrifuge to remove debris prior to analysis. Samples can be stored at -20°C or -80°C. Avoid repeated freeze-thaw cycles.

Preparation Before Use:

Allow samples to reach room temperature prior to assay. Take care to agitate patient samples gently in order to ensure homogeneity.

Test Sample preparation - Samples have to be diluted 1:1000 (v/v), e.g. for 1:1000 (1 ul sample + 999 ul sample diluent) prior to assay. The samples may be kept at 2 - 8°C for up to three days. Long-term storage requires -20°C.

Reagent Preparation (all reagents should be diluted immediately prior to use):

1. Label any aliquots made with the kit Lot No and Expiration date and store it at appropriate conditions mentioned.
2. Bring all reagents to Room Temperature before use.
3. To make **(1X) Wash Buffer**; dilute **25 ml of 20X Wash Buffer in 475 ml of DI water**.
4. **Standards Preparation:** Reconstitute the concentrated Standard lyophilized vial with 1 ml of (1X) Standard Diluent to obtain 4,000 ng/ml. Keep the vial for 15 mins with gentle agitation and then run the assay procedure. Use the (1X) Standard Diluent as the zero standard. Below table shows the calculation for the standard range.

Standard Concentration (ng/ml)	Standard No.	Dilution Particulars
4000 ng/ml	Lyophilized standard	Lyophilized Standard + 250ul of Standard Diluent (1X)
16000 ng/ml	Standard No.5	Reconstituted Standard only
8000 ng/ml	Standard No.4	125 ul Standard No.5 + 125 ul Standard Diluent (1X)
4000 ng/ml	Standard No.3	125 ul Standard No.4 + 125 ul Standard Diluent (1X)
2000 ng/ml	Standard No.2	125 ul Standard No.3 + 125 ul Standard Diluent (1X)
1000 ng/ml	Standard No.1.	125 ul Standard No.2 + 125 ul Standard Diluent (1X)
0 ng/ml	Standard No.0	500 ul Standard Diluent (1X)

Mix each tube thoroughly before the next transfer. Use the standards for experiment within one hour of preparation of standard. Discard standard after use.

5. **Working Ofatumumab :HRP Conjugate – Refer to the Reagent Preparation sheet attached with the IFU and COA (enclosed in the kit).**

Procedural Notes:

1. In order to achieve good assay reproducibility and sensitivity, proper washing of the plates to remove excess un-reacted reagents is essential.
2. Avoid assay of Samples containing Sodium Azide (NaN_3), as it could destroy the HRP activity of the conjugate resulting in under-estimation of the antibodies.
3. It is recommended that all Standards and Samples be assayed in duplicates or triplicates.
4. Maintain a repetitive timing sequence from well to well for all the steps to ensure that the incubation timings are same for each well.
5. If the Substrate has a distinct blue color prior to use it may have been contaminated and use of such substrate can lead to compromise of the sensitivity of the assay.
6. The plates should be read within 30 minutes after adding the Stop Solution.
7. Make a work list in order to identify the location of Standards and Samples.
8. Making serial dilution in the wells directly is not permitted.
9. Prepare the Standard within 15 minutes prior to running the assay.
10. Please carefully dilute Standards according to the instruction, and avoid foaming. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettes are calibrated.
11. If crystals have formed in the Wash Solution (20X) concentrate, warm to room temperature and mix gently until the crystals are completely dissolved.
12. Contaminated water or container for reagent preparation will influence the detection results.

Assay Procedure:

1. It is strongly recommended that all Standards and Samples be run in duplicates or triplicates. A standard curve is required for each assay.
2. Pipette out **100 ul of Standards and samples** to the respective wells.
3. Add **100 ul working Ofatumumab:HRP Conjugate** to each well.
4. Cover the plate and incubate for 90 mins at 37°C.
5. Aspirate and wash plate 4 times with **Wash Buffer (1X)** and blot residual buffer by firmly tapping plate upside down on absorbent paper. Wipe off any liquid from the bottom outside of the microtiter wells as any residue can interfere in the reading step.
6. Add **100 ul of TMB Substrate** in each well.
7. Incubate the plate at RT for 30 minutes in dark. DO NOT SHAKE or else it may result in higher backgrounds and worse precision. Positive wells should turn bluish in color.
8. Pipette out **100 ul of Stop Solution**. Wells should turn from blue to yellow in color.
9. Read the absorbance at 450 nm with a microplate reader.

Calculation of Results:

Determine the Mean Absorbance for each set of duplicate or triplicate Standards and Samples. Using Semi-Log graph paper, plot the average value (absorbance 450nm) of each standard on the Y-axis versus the corresponding concentration of the standards on the X-axis. Draw the best fit curve through the standard points.

To determine the unknown Ofatumumab concentrations, find the unknown's Mean Absorbance value on the Y-axis and draw a horizontal line to the standard curve.

At the point of intersection, draw a vertical line to the X-axis and read the Ofatumumab Concentration.

If samples were diluted, multiply by the appropriate dilution factor. Software which is able to generate a cubic spline curve-fit or 4PL (2nd order) is best recommended for automated results.

Note:


It is recommended to repeat the assay at a different dilution factor in the following cases:

- If the sample absorbance value is below the first standard.
- If the absorbance value is equivalent or higher than the 4000 ng/ml standard.

Typical Data

Standard Concentration (ng/ml)	Mean Absorbance	Interpolated Concentration	% Interpolated Concentration against Actual Concentration
0	2.215	--	--
1000	1.325	1131.7	113.2
2000	1.125	1735.9	86.8
4000	0.413	4853.6	121.3
8000	0.342	7366.6	92.1
16000	0.221	13482.6	84.3

Typical Graph



abs = absorbance at 450nm

Quality Control:

It is recommended that for each laboratory assay appropriate quality control samples in each run to be used to ensure that all reagents and procedures are correct.

Performance Characteristics:

Please note that this validation is performed in our laboratory and will not necessarily be duplicated in your laboratory. This data has been generated to enable the user to get a preview of the assay and the characteristics of the kit and is generic in nature. We recommend that the user performs at the minimum; the spike and recovery assay and the dilutional linearity assay to assure quality results.

For a more comprehensive validation, the user may run the protocols as suggested by us herein below to develop the parameters for quality control to be used with the kit.

Standard Range:

1000 ng/ml - 16000 ng/ml.

Sensitivity:

The minimum detectable dose of Ofatumumab is 990 ng/ml.

Safety Precautions:

- **This kit is For Research Use only.** Follow the working instructions carefully.
- The expiration dates stated on the kit are to be observed. The same relates to the stability stated for reagents
- Do not use or mix reagents from different lots.
- Do not use reagents from other manufacturers.
- Avoid time shift during pipetting of reagents.
- All reagents should be kept in the original shipping container.
- Some of the reagents contain small amount of sodium azide (< 0.1 % w/w) as preservative. They must not be swallowed or allowed to come into contact with skin or mucosa.
- Source materials maybe derived from human body fluids or organs used in the preparation of this kit were tested and found negative for HBsAg and HIV as well as for HCV antibodies. However, no known test guarantees the absence of such viral agents. Therefore, handle all components and all patient samples as if potentially hazardous.
- Since the kit contains potentially hazardous materials, the following precautions should be observed
 - Do not smoke, eat or drink while handling kit material
 - Always use protective gloves
 - Never pipette material by mouth
 - Wipe up spills promptly, washing the affected surface thoroughly with a decontaminant.
- In any case GLP should be applied with all general and individual regulations to the use of this kit.




SCHEMATIC ASSAY PROCEDURE

1. Remove all components, 30 minutes before adding into the assay plate.






2. Avoid repeated cool-thaw of the components as there will be a loss of activity and this can affect the results.





3.  Pipette **100 ul prepared Standard and Sample** into respective well.

4.  Pipette **100 ul working Ofatumumab:HRP Conjugate** into each wells.

5.  Cover plate and incubate for  at 37°C.

6.  Aspirate and wash wells **4 times** with **Wash Buffer (1X)**.

7.  Pipette **100 ul TMB Substrate** into each well.

8.  Cover plate and incubate for  at Room Temperature.

9.  Pipette **100 ul Stop Solution** into each well.

10. Read absorbance at 450nm with a  microplate reader within  of stopping reaction.

Typical Example of a Work List

Well #	Contents	Absorbance at 450nm	Mean Absorbance	ng/ml Ofatumumab
1A	zero standard			
2A	zero standard			
1B	1000 ng/ml			
2B	1000 ng/ml			
1C	2000 ng/ml			
2C	2000 ng/ml			
1D	4000 ng/ml			
2D	4000 ng/ml			
1E	8000 ng/ml			
2E	8000 ng/ml			
1F	16000 ng/ml			
2F	16000 ng/ml			
1G	Sample			
2G	Sample			
1H	Sample			
2H	Sample			

LIMITED WARRANTY

Krishgen Biosystems does not warrant against damages or defects arising in shipping or handling, or out of accident or improper or abnormal use of the product; against defects in products or components not manufactured by Krishgen Biosystems, or against damages resulting from such non-Krishgen Biosystems made products or components. Krishgen Biosystems passes on to customer the warranty it received (if any) from the maker thereof of such non-Krishgen made products or components. This warranty also does not apply to product to which changes or modifications have been made or attempted by persons other than pursuant to written authorization by Krishgen Biosystems.

THIS WARRANTY IS EXCLUSIVE. The sole and exclusive obligation of Krishgen Biosystems shall be to repair or replace the defective product in the manner and for the period provided above. Krishgen Biosystems shall not have any other obligation with respect to the products or any part thereof, whether based on contract, tort, strict liability or otherwise. Under no circumstances, whether based on this Limited Warranty or otherwise, shall Krishgen Biosystems be liable for incidental, special, or consequential damages.

This Limited Warranty states the entire obligation of Krishgen Biosystems with respect to the product. If any part of this Limited Warranty is determined to be void or illegal, the remainder shall remain in full force and effect.














Krishgen Biosystems. 2025

THANK YOU FOR USING KRISHGEN PRODUCT!

KRISHGEN BIOSYSTEMS®, GENLISA®, DHARMAPLEX™, GENBULK™, GENLISA™, KRISHZYME®, KRISHGEN®, KRIBIOLISA®, KRISHPLEX®, TITANIUM®, QUALICHEK® are registered trademarks of KRISHGEN BIOSYSTEMS. ©KRISHGEN BIOSYSTEMS. ALL RIGHTS RESERVED.

KRISHGEN BIOSYSTEMS | OUR REAGENTS | YOUR RESEARCH |

SYMBOLS KEY

	CLAUDIN-18 protein Coated Microtiter Plate (12x8 wells)
	Ofatumumab Standard
	Ofatumumab:HRP Conjugate concentrated
	Detection Diluent
	(1X) Standard Diluent
	(1X) Sample Diluent
	(20X) Wash Buffer
	TMB Substrate
	Stop Solution
	Consult Instructions for Use
	Catalog Number
	Expiration Date
	Storage Temperature